



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

FILED

JUL 20 2006

KCDC EAST DIVISION
ORELINE COURTHOUSE

DATE: 4/28/2005 INITIALS: DEC OWNER: WSP
DISTRICT: 2 ADDRESS: I-5 So MP 141
LOCATION: Septac No 27 COUNTY: King
MAKE: Volvo TYPE: truck scale ACCURACY CLASS: III
MODEL: 8530 ID. NO: 4338198-4NV CAPACITY: 100,000 SECTIONS: 2
MINIMUM GRADUATIONS: 20 lb REG. ELEMENT: digital PLATFORM: 12' X 20'


| LOAD POSITION | TEST WEIGHT | SCALE INDICATION | ERROR | LOAD POSITION | TEST WEIGHT | SCALE INDICATION | ERROR |
|---------------|---------------|------------------|------------|---------------|-------------|------------------|-------|
| <u>Sec 1</u> | <u>20,000</u> | <u>19,980</u> | <u>-20</u> | | | | |
| <u>Center</u> | <u>20,000</u> | <u>20,000</u> | <u>0</u> | | | | |
| <u>Sec 2</u> | <u>20,000</u> | <u>20,000</u> | <u>0</u> | | | | |
| | | | | | | | |
| | | | | | | | |

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as is shown on the attached "Certificate of Accuracy", No. L 2204 - 7 and 8

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Crona Date: 5/23/2005

State of Washington
County of Thurston
Signed and sworn to (or affirmed) before
me on May 23 2005 by
Michael J. Evans
(Signature of notary public)
Michael J. EVANS
Printed Name of Notary
My appointment expires 12-27-07



3000-362-021 (R) (10/98)

DISTRIBUTION: WHITE - COURT GREEN - DISTRICT CANARY - PROPERTY MANAGEMENT PINK & GOLDENROD - SERVICE TRUCKS